

University College Cork, Ireland Coláiste na hOllscoile Corcaigh

AUTHOR QUERY FORM

Dear Author,

Below are the queries associated with your article. Please answer all of these queries before sending the proof back to AIP.

Thank you for your assistance.

¹ THE JOURNAL OF CHEMICAL PHYSICS **137**, 000000 (2012)

[Hollow core photonic crystal fiber based viscometer](http://dx.doi.org/10.1063/1.4771659) [with Raman spectroscopy](http://dx.doi.org/10.1063/1.4771659) 2 3

L. E. Horan, $1,2, a$) A. A. Ruth, $2, b$) and F. C. Garcia Gunning $1,2, c$)

¹ ⁵ *Tyndall National Institute, University College Cork, Lee Maltings, Cork, Ireland*

² ⁶ *Department of Physics, University College Cork, Lee Maltings, Cork, Ireland*

⁷ (Received 27 September 2012; accepted 28 November 2012; published online XX XX XXXX)

The velocity of a liquid flowing through the core of a hollow core photonic crystal fiber (driven by capillary forces) is used for the determination of a liquid's viscosity, using volumes of less than 10 nl. The simple optical technique used is based on the change in propagation characteristics of the fiber as it fills with the liquid of interest via capillary action, monitored by a laser source. Furthermore, the liquid filled hollow core photonic crystal fiber is then used as a vessel to collect Raman scattering from the sample to determine the molecular fingerprint of the liquid under study. This approach has a wide variety of indicative uses in cases where nano-liter samples are necessary. We use 10–12 cm lengths of hollow core photonic crystal fibers to determine the viscosity and Raman spectra of small volumes of two types of monosaccharides diluted in a phosphate buffer solution to demonstrate the principle. The observed Raman signal is strongest when only the core of the hollow core photonic crystal fiber is filled, and gradually decays as the rest of the fiber fills with the sample. *© 2012* 8 9 10 11 12 13 14 15 16 17 18

American Institute of Physics. [\[http://dx.doi.org/10.1063/1.4771659\]](http://dx.doi.org/10.1063/1.4771659) 19

²⁰ **I. INTRODUCTION**

 $_{21}$ Hollow core photonic crystal fibers (HC-PCFs)^{1–[3](#page-8-1)} have attracted a lot of interest due to the possibility of inserting ²³ liquids into the capillaries for optical sensing in biological^{4,[5](#page-8-3)} and chemical applications.⁶ These applications tend to exploit its main optical property called the photonic bandgap (PBG) ϵ effect,^{[7,](#page-8-5)[8](#page-9-1)} where light guidance occurs within the hollow core only for certain wavelengths. Another interesting effect in such fibers, also used in sensing, is the bandgap shift that α occurs when all capillaries are filled with a liquid sample.⁹ In this case, light propagation occurs at significantly shorter 31 wavelengths than the original PBG. The shift in PBG can be α estimated utilizing scalar wave equations.⁸ Finally, if only the core of the fiber is selectively filled, the optical propagation once again changes, with the physical mechanism changing σ from PBG to index-like guidance.¹⁰ In this paper, we exploit all these changes in propagation properties during the inser- tion of a liquid into the fiber in order to determine its dynamic viscosity by its flow through the hollow capillaries. The flow of liquids through ordinary capillaries has been extensively 40 studied and is widely used to determine viscosities.^{[11](#page-9-4)} Com- pared with current commercially available capillary viscome- ters (1–10 ml volume samples required), the proposed tech- nique here requires only nanoliter quantities of liquid for any measurement.

⁴⁵ The technique shown here was tested using phosphate ⁴⁶ buffer saline (PBS) solutions of the monosaccharides glu-⁴⁷ cose and fructose across a range of concentrations. Monosac-48 charides are ubiquitous in life and food sciences, 12 12 12 e.g., as ⁴⁹ a source of energy in living organisms and as a fundamental ingredient in food for the enhancement of flavour or as a 50 preservative. Knowing the viscosity of saccharide solutions, 51 for example, is important for the measurement of the degree 52 of sweetness.^{[13](#page-9-6)} Here we will also show that, with a slight $\overline{}$ s modification, it is possible to characterize the composition of $_{54}$ the liquid sample via the detection of vibrational Raman scat- ⁵⁵ tering within the hollow core. HC-PCF has been shown to be 56 an excellent medium for the collection of Raman scatter^{[14,](#page-9-7) [15](#page-9-8)} $\frac{57}{2}$ when it is filled with solutions for two reasons: (a) the resultant wavelength shift with a broad PBG window enables the 59 guiding of both the excitation light and the Raman scatter at \approx visible wavelengths through the fiber directly to the detector, 61 leading to high collection efficiencies; and (b) the possibility ϵ ₈₂ of filling long lengths of fibers enhances the light-liquid in- ⁶³ teraction path length and increases the measurement sensitiv- 64 ity. Therefore, this viscometer system can provide informa- ⁶⁵ tion about the molecular identity of the solution in question $\overline{66}$ without extensive further analysis. 67

II. NANOLITER VISCOMETER 68

A. Principle 69

In this paper, we use a simple force model derived 70 from Nielsen *et al.*,^{[16](#page-9-9)} which is based on the Lucas-Washburn $\frac{1}{71}$ equation.^{[17](#page-9-10)} The Lucas-Washburn equation describes the time $\frac{72}{2}$ it takes for a liquid to flow through a capillary. This allows us 73 to determine the time it takes to fill a certain length of HC- ⁷⁴ PCF, and thus the average velocity of the liquid flow. This $\frac{75}{6}$ model takes into account all forces that affect capillary flow: $\frac{76}{6}$ the capillary and overpressure forces, which are balanced by 77 frictional and gravitational forces. In our experiments, over- ⁷⁸ pressure was not applied and gravity can be neglected due to ⁷⁹ the small core radius of the HC-PCF (typical radius ∼5 *μ*m). ⁸⁰
We ignore the initial condition of zero velocity, which should ⁸¹ We ignore the initial condition of zero velocity, which should

a[\)laura.horan@tyndall.ie.](mailto: laura.horan@tyndall.ie)

b)_{a.ruth@ucc.ie.}

 $c)$ fatima.gunning@tyndall.ie.

82 produce negligible error due to the length of fiber used here.

83 The model also assumes laminar flow and that the contact an-84 gle with the capillary walls remains constant throughout the ⁸⁵ filling process. This is a reasonable assumption, given the en-⁸⁶ vironmental conditions are stable and only Newtonian liquids 87 are considered. Thus, the average flow velocity (*v*) will de-88 pend on the liquid's surface tension (σ) and viscosity (μ) and ⁸⁹ on the capillary radius (*r* ∼ 5 *μ*m),

$$
v = \left(\frac{r}{2L}\right) \left(\frac{\sigma}{\mu}\right) = \frac{L}{t},\tag{1}
$$

⁹⁰ and is simply given by the ratio of fiber length (*L* ranging ⁹¹ from 10–12 cm) to filling time (*t*). This relationship is unique $\frac{18}{18}$ $\frac{18}{18}$ $\frac{18}{18}$ to each liquid.¹⁸ When the liquid's surface tension, the fiber ⁹³ length, and capillary radius are known, and the filling time is 94 measured, the viscosity can be determined using Eq. [\(1\).](#page-3-0)

⁹⁵ In order to determine the surface tension of a solution ⁹⁶ with a given solute concentration of *C*, a well-known linear 97 approximation utilizing static parameters may be used:¹⁹

$$
\sigma = \left(\sigma_w + \frac{\Delta \sigma}{\Delta C} C\right) \cos \theta, \tag{2}
$$

98 where σ_w is the surface tension of the solvent (PBS in our es case), with a constant $\Delta \sigma / \Delta C = 1.3 \times 10^{-3}$ (N m⁻¹)/M¹⁹ 100 for both glucose and fructose, and cos θ is the angle that the ¹⁰¹ meniscus makes with the capillary walls. To determine the 102 surface tension $σ$, the system is calibrated with the solvent $_{103}$ only ($C = 0$). This is done by measuring the average flow ve-¹⁰⁴ locity of the solvent PBS, and calculating the surface tension σ in Eq. [\(1\),](#page-3-0) using the viscosity value from the literature.^{[20](#page-9-13)} 106 The value found for σ is then inserted into Eq. [\(2\)](#page-3-1) to deter-107 mine the cosine of the contact angle, taking $\sigma_w = 0.0695$ N/m $f₁₀₈$ from literature.^{[21](#page-9-14)} In this experiment, the angle *θ* was esti- $\frac{109}{109}$ mated to be \sim 50°, using our results and known values for ¹¹⁰ PBS. The viscosities of the saccharide solutions outlined in 111 Sec. [II D](#page-5-0) were then determined with Eq. (1) using the surface 112 tension from Eq. [\(2\)](#page-3-1) and the measured average velocity of the ¹¹³ fluid.

¹¹⁴ **B. Experimental setup**

 The fiber chosen was HC-PCF-1060 (NKT Photonics A/S), with an original primary PBG at 1060 nm, a secondary PBG at 735 nm, and a hollow core radius of ∼5 *μ*m. The fiber core had a volume of 9.4 nl for a length of 12 cm. The total 119 cross-sectional radius of the microstructure was $25 \mu m$. For all experiments in this paper, the fiber segments were cleaved 121 and not subjected to chemical pre-treatment.

 The experimental setup used to determine the aver- age velocity is schematically illustrated in Fig. [1.](#page-3-2) Two laser sources were used. The first was a supercontinuum source (SC) used for alignment of the empty fiber us- ing a bandpass filter at 1050 nm (with 10 nm FWHM) and subsequently for characterizing the light propagation changes during fiber filling. This is discussed in more de- tail in Sec. [II C](#page-4-0) below. The SC source consisted of a semiconductor-pumped Q-switched Nd:YAG laser at 1064 131 nm (6.85 kHz repetition rate, \sim 0.55 ns pulses, 73.5 mW av-erage power), which pumped a highly nonlinear fiber (SC

FIG. 1. Schematic of the viscometer experimental setup. A supercontinuum source (SC) was used to align the fiber with the aid of a \times 40 objective lens (L) to focus the light to the fiber. A flip mirror was used to guide 633 nm light during the fluid flow within the fiber. The output signal was collimated by a \times 10 objective lens, and divided by a beam splitter (BS) to enable light monitoring by a photodiode (PD) and a charged coupled detector (CCD).

5 1040, Blaze Photonics) and generated a broadband emis- ¹³³ sion over the wavelength range (∼400–1750 nm). The sec- 134 ond source was a continuous wave (cw) HeNe laser (\sim 1 mW, 13563 nm) used in order to determine the average velocity of the 136 633 nm) used in order to determine the average velocity of the ¹³⁶ liquid flowing through the fiber. In order to maintain the initial alignment, a flip-mirror was inserted in the setup to easily 138 switch between optical sources. Light at 633 nm should ini- ¹³⁹ tially experience high-loss because it falls outside the PBG of ¹⁴⁰ the empty fiber. However, once the fiber is filled with a liquid $_{141}$ of refractive index ∼1.33, the predicted PBG shift⁹ would be 142
centered at ~600 nm with a bandwidth of ~200 nm, enabling 143 centered at \sim 600 nm with a bandwidth of \sim 200 nm, enabling low-loss transmission of the HeNe laser light.

The collimated laser light was focused into the HC-PCF 145 by a \times 40 objective lens. Fiber segments of 10–12 cm in length 146 were chosen, which were long enough to enable accurate optical alignment, and yet short enough to limit the measure- ¹⁴⁸ ment time and to prevent excessive temperature variations. ¹⁴⁹ The fiber tips had their cladding jacket removed, 2 cm from 150 each end, in order to allow cleaving of the ends for better light 151 coupling. 152

Viscosity is strongly temperature dependent, and gener- ¹⁵³ ally increases exponentially with decreasing temperature. 22 154 As such, steps were taken to ensure the fiber sample was 155 kept at constant temperature along its length during the mea- ¹⁵⁶ surement. The schematic of our temperature control system is 157 shown in Fig. [2.](#page-3-3) The *xyz* optical positioning stage (not shown 158) in Fig. [2\)](#page-3-3) holding the fiber was fitted with a 5-cm-long cop- ¹⁵⁹ per casing which kept the temperature of the fiber at 29.0 160 \pm 0.1 °C with the aid of a thermoelectric Peltier cooler (TEC). 161 The temperature was monitored by a thermistor positioned 162

FIG. 2. Schematic for fiber holder and temperature control system for the HC-PCF (not to scale). Liquid is inserted into the reservoir on the right and the flow inside the fiber occurs from right to left of the figure. The laser light is launched at the left side of the figure, and detected by the PD and CCD at the right side via the PMMA window.

FIG. 3. Cross-sectional images of a HC-PCF-1060 showing its dynamic filling (a) – (e) observed via an optical microscope. The water reservoir, in this case, was placed at the other end of the fiber. The images show that the core fills faster (b) than the outer capillaries due to the difference in capillary radius.

 just above the groove where the fiber was held in place. The fiber then exited the copper enclosure and entered the empty liquid reservoir through a series of plastic connectors in order to avoid leakage of the liquid when the reservoir is later filled. The reservoir had a small polymethyl methacrylate (PMMA) window on the detector side in order to allow the transmitted light to be collected. The reservoir was also maintained at a constant temperature, with an additional TEC and thermistor monitor placed at the side of the copper reservoir. Tempera- ture could not be controlled within the plastic connectors, but this part of the fiber was kept to a minimum. These measures helped to minimize temperature gradients along the HC-PCF ¹⁷⁵ length.

 To detect the propagation changes, the output of the fiber was collimated with a $\times 10$ infinity corrected lens (Fig. [1\)](#page-3-2) to monitor the power of the light transmitted by the fiber. A beam splitter was added to facilitate initial alignment and near field monitoring with a charged coupled detector (CCD) camera (color sensor Sony ICX204AK). The signal from the low $s₁₈₂$ speed photodiode (PD) (Silicon, active area 13 mm², 350– 1100 nm, cathode grounded) was recorded with an analog-to- digital converter connected to a computer. The CCD images were recorded via USB at 30 frames/s.

¹⁸⁶ **C. Viscosity results**

¹⁸⁷ An optical microscope (Nikon Eclipse ME600) was used ¹⁸⁸ to understand the filling process in real time, as shown in Fig. [3.](#page-4-1) The HC-PCF-1060 was filled by water at one end and 189 observed by an optical microscope with a magnification of ¹⁹⁰ \times 50 at the other. The images initially show the empty fiber 191 (Fig. $3(a)$), the point in time where only the core is filled 192 (Fig. $3(b)$), a sequence of gradual filling of the capillaries 193 (Figs. $3(c)$ and $3(d)$), and finally when the entire fiber is fully 194 filled (Fig. $3(e)$). The core fills faster than the capillaries due 195 to its larger diameter, as expected by Eq. (1) .

As a consequence, the light propagation during the filling 197 process will also dynamically change, and different ranges of 198 wavelengths will propagate at the different stages exemplified in Fig. [3.](#page-4-1) To illustrate the wavelength dependent propa- ²⁰⁰ gation features, a selection of near-field color-sensitive CCD ²⁰¹ images was taken during the filling process and shown in ²⁰² Fig. [4](#page-4-2) for a fiber sample of 10 cm. For that purpose, the 203 broadband emission from the SC source was used to measure ²⁰⁴ the light transmission during the filling. Initially $(t = 0 s)$, 205 only light at wavelengths within the original primary PBG ²⁰⁶ centered around 1060 nm and secondary PBG are guided, ²⁰⁷ as the reservoir is empty (Fig. $4(b)$). Note that near-infrared 208 wavelengths are detected by the CCD and are shown in ²⁰⁹ Fig. $4(b)$ as an intense violet color. With increasing time, 210 the filling fiber displays several guidance mechanisms. First, ²¹¹ guidance by PBG effect dominates initially as the fiber is ²¹² empty, and then changes to index guiding (once the core is 213 fully filled), and finally to wavelength-shifted PBG (when all 214 capillaries are filled). In more detail, while the fiber is ini- ²¹⁵ tially filling, the HC-PCF segment at the laser side of the fiber ²¹⁶ is still empty, therefore the original PBG dominates and light 217 at wavelengths outside the PBG bandgap are scattered. On ²¹⁸ the detector side of the fiber, however, where the fluid is fill- ²¹⁹ ing the capillaries, some of the scattered light is able to cou- ²²⁰ ple back into the core either by index-like guiding or by the ²²¹ PBG shift (Fig. $4(c)$). Eventually, after 114 s (Fig. $4(d)$), the 222 core is completely filled, the image from the CCD is satu- ²²³ rated as an index-like guiding mechanism allows for a broad ²²⁴ range of visible wavelengths to be propagated. After 340 s 225 (not shown in Fig. [4\)](#page-4-2), all capillaries are filled, and a PBG ²²⁶ shift was observed. Evidence of this is shown in Figs. $4(e)$ 227 and $4(f)$, when bandpass filters (FWHM 10 nm) centered at $_{228}$

FIG. 4. (a) Image of the cross section of the HC-PCF-1060 fiber segment used taken by a Nikon Eclipse ME600 optical microscope. (b)–(d). Near-field images of the fiber transmitting broadband light from a supercontinuum source at different stages during the filling process. (d) Image of the fiber at 114 s when the core is filled. Panels (e) and (f) show pictures of the fiber where the light is transmitted through bandpass filters centered at 550 nm and 650 nm, respectively. Note: The CCD camera is light sensitive between 400 nm and 1100 nm. (g) Transmission spectra of 10 cm fiber segment when empty (blue triangles), core-only filled (purple squares), and fully filled (green circles). Also shown is the emission spectrum of the supercontinuum waveband (black diamonds.) Insets are the corresponding images depicting the filling states of the fiber in conjunction with the spectral outputs of the fiber for each filling state.

²²⁹ 550 nm and 650 nm, respectively, were inserted to confirm ²³⁰ the broad range of wavelengths guided between ∼500 and ²³¹ 800 nm.

 The output spectrum of the HC-PCF was measured using an optical spectrum analyser (OSA), when the SC source was launched to the fiber, as in Fig. [1.](#page-3-2) The results are presented in Fig. $4(g)$, exemplifying the changes in guidance properties for the three different filling configurations: empty, core-only filled, and fully filled. The blue spectrum (triangles) shows 238 guidance by the original PBG (Fig. $4(b)$), with a maximum at 1060 nm, and a secondary band at ∼735 nm. The purple spec- trum (squares) was recorded when only the core was filled, and where index guiding is the dominant guiding mechanism (Fig. [4\(d\)\)](#page-4-2). This transmission exhibits two distinct peaks: the first, at 1060 nm, relates to the original PBG determined by the cladding capillaries structure, and a broadband guidance centered at ∼810 nm, due to index guiding effect. This broad- band guidance had 3 dB (equivalent to a full width half max- imum) and 20 dB bandwidths of at least 150 nm and 440 nm, respectively, and the spectrum is clipped at lower wave- lengths due to the non-optimal condition of the supercontin- uum source used at the time, as observed in the black spec- trum (diamonds). The spectrum represented by green circles was measured when the entire fiber was filled with water. The 253 spectrum features a shifted PBG at \sim 750 nm (Figs. [4\(e\)](#page-4-2) and $254 \frac{4(f)}{h}$, with a 3 dB bandwidth of at least 100 nm and a 20 dB bandwidth of 280 nm, respectively. The response of our SC source at lower wavelengths needs to be taken into consider-ation when analysing the PBG shift for the fully filled fiber.

 The observations from Fig. [4](#page-4-2) tell us that, with careful selection of the hollow core fiber and light sources, optical propagation can be either suppressed or supported depending on the filling state. Based on this principle, it is reasonable to expect that the use of a narrow-band source (in our case, a HeNe laser at 633 nm) will allow monitoring of the filling evolution of the core and the capillaries of a HC-1060 fiber, and hence determine the viscosity properties of the respective sample fluid.

²⁶⁷ Figure [5](#page-5-1) shows the results obtained using the experi-²⁶⁸ mental setup from Fig. [1.](#page-3-2) After fiber alignment with the

FIG. 5. HC-1060 fiber transmission at 633 nm as a function of time during the filling process. The fiber was filled with a 0.5 M solution of glucose in water in this example. The transmission is represented by the signal of the PD in units of volts. Insets show the near-field images of the output of the HC-PCF, captured by the CCD camera at different times.

SC source, light coupling optimization with the HeNe laser 269 source, and temperature stabilisation, the aqueous solution 270 was inserted into the reservoir. Figure [5](#page-5-1) shows the fiber trans- 271 mission as a function of time, detected by the PD. Although 272 no-guidance at the core was expected for the empty fiber at t_{273} $= 0$, a signal was still observed. We believe this is due to a $_{274}$ small amount of scattered light from the core becoming consmall amount of scattered light from the core becoming confined in the solid cladding. Inserting the aqueous solution into 276 the reservoir caused a sudden drop in the PD signal at \sim 10 s. 277 The initially decreasing PD signal in Fig. 5 was caused by 278 changes in the output alignment of the fiber to the PD, due ²⁷⁹ to refraction of the exiting beam when the liquid was inserted 280 into the reservoir. However, this sudden drop also defined the ²⁸¹ start time t_0 for the filling process, and from here onwards the 282 coupling remained constant. As the capillaries begin to fill, ²⁸³ the change in the propagation properties is apparent. A rapid ²⁸⁴ increase in transmission was observed just before *t* ∼ 200s, ²⁸⁵ which we believe coincides with the onset of index-guided 286 propagation resulting from the filling of the fiber core. Once 287 the core was fully filled after $t \sim 185$ s, the transmission re- 288 mained almost constant.

A measurement of the core filling time is sufficient to ²⁹⁰ determine the average velocity for the viscosity analysis as ²⁹¹ described in Sec. [II A.](#page-2-3) Nevertheless, as the filling continued ²⁹² through the cladding capillaries, modal competition within 293 the core was observed until the capillaries were completely ²⁹⁴ filled after 900 s. The insets in Fig. 5 show the propagation 295 changes captured by the CCD camera at different times dur- ²⁹⁶ ing the filling.

D. Analyses for glucose and fructose solutions ²⁹⁸ **in PBS** ²⁹⁹

To demonstrate the principles of operation of the pro- ³⁰⁰ posed viscometer, phosphate buffer saline (PBS, purchased 301 in tablet form from Sigma-Aldrich) was prepared by dissolv- ³⁰² ing one PBS tablet in 200 ml of purified water to obtain ³⁰³ 137 mM NaCl, 2.7 mM KCl, and 10 mM phosphate buffer 304 solution. Solutions of D-(+)-glucose, \geq 99.5% in PBS and D- 305 (−)-fructose, ≥99% (Sigma-Aldrich) in PBS were prepared ³⁰⁶ at concentrations ranging from 10^{-4} M to 1 M. 307

The average filling velocity $(\langle v \rangle)$ was measured for these \sim 308 solutions as a function of concentration.^{[23](#page-9-16)} In Fig. $6(a)$, red 309 diamonds (closed symbol) and blue squares (open symbol) ³¹⁰ represent the results for glucose and fructose solutions, re- ³¹¹ spectively. For concentrations below 0.01 M, $\langle v \rangle$ appears to 312 be constant with an average value of (1.1 ± 0.2) mm s⁻¹. 313 The error was estimated from the uncertainty in measuring 314 length and filling times of the fiber. The random variations of 315 $\langle v \rangle$ for concentrations below 0.01 M are probably due to tem- $\frac{316}{2}$ perature fluctuations between measurements or temperature 317 gradients within the setup. Above this concentration level, ³¹⁸ surface tension 19 and viscosity values increase. There is a 319 notable decrease in the average velocity due to the decrease ³²⁰ in the ratio of surface tension to viscosity. The experimen- ³²¹ tal data in Fig. $6(a)$ were plotted and compared with a con- 322 centration dependant model based on Eq. [\(2\)](#page-3-1) and a well- ³²³ known relationship between viscosity and concentration from 324

FIG. 6. Results obtained with the proposed viscometer. (a) Experimental average velocities for PBS solutions of glucose (red, closed diamond symbols) and fructose (blue, open squash symbols) for different concentrations at 29 °C. Red dashed line and blue dotted line are a ratio of Eq. [\(2\)](#page-3-1) to Eq. [\(3\),](#page-6-1) determined from literature.^{20, [23,](#page-9-16) [27](#page-9-17)[–29](#page-9-18)} (b) Measured viscosity from $\langle v \rangle$ with concentration for the glucose (red, closed diamond symbols) and fructose (blue, open squash symbols) in PBS solutions. The red dashed and blue dotted lines are fitted using Eq. [\(3\)](#page-6-1) and data from the literature, $20,27-29$ $20,27-29$ also presented in Table [I.](#page-6-2)

325 Jones *et al.*^{[24](#page-9-19)}

$$
\mu(C) = \mu_w + A\sqrt{C},\tag{3}
$$

 ω ₂₂₆ where μ_w is the viscosity of the solvent and *A* is a param- 327 eter that depends on the solute. The value for μ_w is taken ³²⁸ from literature to be 1.002 mPa s^{20} s^{20} s^{20} and *A* is calculated from ³²⁹ literature^{20, [26–](#page-9-20)1} $\left\{\sum\right\}$ fitting Eq. [\(3\)](#page-6-1) to the values outlined in 330 Table [I.](#page-6-2) For glucose, $A = 6.08 \times 10^{-4}$ (Pa s)/(M^{1/2}) and for f_{331} fructose, $A = 6.01 \times 10^{-4}$ (Pa s)/(M^{1/2}). The ratio of Eq. [\(2\)](#page-3-1) to $_{332}$ Eq. [\(3\)](#page-6-1) (σ (*C*)/ μ (*C*)) times a constant (r/2L) is then compared 333 to the data acquired for the velocity of liquid flow in Fig. $6(a)$ ³³⁴ for glucose (dashed red line) and fructose (dotted blue line) ³³⁵ solutions.

 $\frac{336}{25,26}$ $\frac{336}{25,26}$ $\frac{336}{25,26}$ The results for glucose and fructose are very similar, $\frac{25,26}{25}$ 337 as expected due to their similar physical parameters.

TABLE I. Viscosity values from literature for saccharide solutions in water.

Concentration (M)	Viscosity glucose (mPa s)	Viscosity fructose (mPa s)	Reference
0.028	1.01	1.01	Mathpal et al. ²⁷
0.056	1.02	1.02	
0.084	1.03	1.04	
0.112	1.05	1.05	
0.14	1.06	1.06	
0.168	1.08	1.08	
0.196	1.09	1.09	
0.224	1.11	1.10	
0.252	1.12	1.12	
0.28	1.14	1.13	
0.308	1.16	1.14	
0.336	1.17	1.16	
$\overline{0}$	1.002		Comesaña et al. ²⁰
0.5	1.26		
1	1.59		
1.5	1.99		
0.27	1.36	1.26	Migliori et al. ²⁸
0.56	1.58	1.43	
1	2.13	2.09	
0.56	1.07	1.02	Telis et al. 29
1	1.68	1.59	

The concentration-dependent viscosity was determined with 338 Eq. [\(1\),](#page-3-0) using surface tension values from Auman *et al.*^{[19](#page-9-12)} and ₃₃₉ Eq. [\(2\)](#page-3-1) as described in Sec. [II A.](#page-2-3) The results are shown in $\frac{340}{2}$ Fig. $6(b)$, in comparison with a selection of data from the $\frac{341}{241}$ literature, $20, 27-29$ $20, 27-29$ $20, 27-29$ which are represented by open symbols and 342 outlined explicitly in Table [I.](#page-6-2) The viscosity measurements ³⁴³ from the literature were determined at similar temperatures 344 in aqueous solutions, but different experimental approaches ³⁴⁵ were utilized. Looking at the trends in Fig. $6(b)$, for low con- 346 centrations the changes in viscosity are expected to be within ³⁴⁷ the error limit, as the amount of saccharide added to the sol- ³⁴⁸ vent is very small. For higher concentrations, however, the ³⁴⁹ viscosity becomes sensitive to the solute concentration. ³⁵⁰ Figure $6(b)$ illustrates this dependence of the viscosity with 351 increasing saccharide concentrations above 0.1 M, showing 352 that the technique proposed here gives similar results than ³⁵³ other well established techniques, with the advantages of ³⁵⁴ small volumes being handled. 355

Although only saccharide solutions were used as test 356 cases, it is evident that this is an effective technique for ana- ³⁵⁷ lyzing the viscosity of liquids in a simple optical setup, which 358 could be miniaturized for a wide range of applications. 359

III. RAMAN SPECTROSCOPY USING HC-PCF 360

Small modifications to our setup can be implemented to 361 allow Raman scattering measurements to be taken in real time 362 as the HC-PCF fills, or after the filling process when the fluid 363 is stable within all capillaries. The detection of Raman scatter 364 from a sample mixture is a powerful analytical tool to identify ³⁶⁵ individual molecular species within a sample. It is known that 366 HC-PCFs can be used to increase the light-sample interaction 367 path length and at the same time improve the light collec- ³⁶⁸ tion efficiency of Raman scattering.^{[15,](#page-9-8) [30](#page-9-23)} Here we show that $\frac{369}{269}$ the Raman scattering can be incorporated with the viscometer 370 technique, enabling dynamic analysis of the fluid while fill- ³⁷¹ ing the capillaries. From our results, we identify that the opti- ³⁷² mum time to take a Raman signal is when the core of the fiber 373 alone is filled with the liquid sample. During the subsequent 374 filling of the capillaries in the micro structured cladding, the 375

FIG. 7. Experimental setup for backscattered Raman experiments, after alignment as outlined in Sec. Π B. Laser light from a frequency doubled cw Nd:YAG laser was launched to the HC-PCF via a beam splitter (BS) and a \times 40 lens (L). While the transmitted signal was monitored via a CCD camera and a PD, as in Fig. [1,](#page-3-2) the backscattering was collected via a lens, a notch filter (NF) to block the excitation beam, and cooled Andor spectrometer.

³⁷⁶ inelastic scattering signal decreases by typically 40%-60% 377 when the entire fiber is filled.

³⁷⁸ **A. Experimental setup**

 A frequency doubled continuous wave (cw) Nd:YAG laser (∼10 mW at 532 nm) was used to optically pump the sample. The Raman scatter was detected with a backscatter onfiguration, as shown in Fig. [7.](#page-7-0) Raman backscatter was col- lected using a \times 40 infinity correction lens, which was also used to focus the light into the fiber initially. The backscat- tered light passed through a beam splitter and a notch filter 386 centered at 532 \pm 17 nm (suppression 10⁻⁶) to block light at the excitation wavelength. The inelastic (Raman) backscat- ter was focused onto a light guide with a lens. The light guide (fiber bundle with round entrance aperture and "slit" exit aper- ture) was connected to a polychromator with back-illuminated and cooled CCD detector (Andor DV401, −30 °C) with a res- olution of 0.2 nm. The output of the fiber in the forward scat- tering regime is continuously monitored by a CCD camera and PD, as detailed in Sec. [II B,](#page-3-4) and shown in Fig. [1](#page-3-2) (detector ³⁹⁵ side).

 Precise control of the light guidance was crucial, and the 397 positioning of the lenses, beam splitter, and the HC-PCF seg- ment in the setup was adjusted to maximize the collection of the Raman signal. Backscattered Raman signal from the silica cladding can decrease the sensitivity of our measurement, and slight optical misalignments will lead to a significant decrease 401 of the signal-to-noise ratio. 402

B. Raman spectra of saccharide solutions

The setup was used to identify the well-known spectra 404 for solutions of glucose and fructose diluted in PBS inside ⁴⁰⁵ a fully filled HC-PCF, shown in red and blue in Fig. [8,](#page-7-1) re- ⁴⁰⁶ spectively, and for the wavenumber range between 290 and 407 3900 cm^{-1} with an integration time 60 s. To obtain the spectra 408 in Fig. 8 , a background spectrum for a fiber filled with solvent 409 PBS only was subtracted, accounting for any scattering from 410 the fiber itself and PBS. The light guiding properties of the 411 fiber may cause the resulting Raman scatter intensities to dif- ⁴¹² fer slightly to those acquired from a bulk sample. However, ⁴¹³ the spectral region addressed comprises the characteristic ⁴¹⁴ C–O and C–C stretch vibrations at 900, 1100, and 1350 cm⁻¹ $_{415}$ for glucose and 600, 1230, and 1430 cm⁻¹ for fructose, al- 416 lowing the identification of the two monosaccharides. $\frac{31}{417}$ $\frac{31}{417}$ $\frac{31}{417}$ 417

C. Simultaneous viscometer and Raman analysis ⁴¹⁸

To demonstrate the feasibility of integrating the vis- ⁴¹⁹ cometer with the Raman scattering setup, a kinetic series of ⁴²⁰ Raman spectra was collected during the filling of a 12 cm ⁴²¹ HC-PCF with 1 M glucose solution. The same experimental 422 procedure described in Sec. [II B](#page-3-4) was applied, by exchanging 423 the HeNe laser (centered at 633 nm) with a cw frequency- ⁴²⁴ doubled Nd:YAG laser at 532 nm. A series of 800 Raman ⁴²⁵ backscatter measurements was taken, as the fiber was filled, ⁴²⁶ each measurement having an integration time of 1s, with 427 the wavenumber shift measured sequentially over a range of 428 4000 cm⁻¹, and a selection of results is plotted in Fig. $9(a)$. 429 Initially no backscattering was observed until 350 s. This time 430 corresponds to the core of the fiber being filled by ∼90%. The ⁴³¹ signal strength reached a maximum at \sim 360 s, when the core 432 was fully filled, enabling propagation of broadband light, but 433 it decayed again after that.

In order to understand the dynamics of the Raman scat- ⁴³⁵ tering signal strength, one must look closely to the optical ⁴³⁶ propagation properties, as discussed in Sec. [II C.](#page-4-0) Along with $_{437}$ the backscattering, the optical power through the fiber was ⁴³⁸

FIG. 8. Stokes Raman spectra for 1M (a) glucose (red) and (b) fructose (blue) in PBS solutions contained within a HC-PCF. Integration times for samples in the HC-PCF was $60 s³$

FIG. 9. Results for an integrated viscometer and Raman scattering system using a 12 cm HC-PCF for a 1 M glucose solution in PBS. (a) Dynamic Raman backscatter taken every 1 s (integration time). (b) Transmission results for photodiode signal (black) during filling, normalised Raman signal at 3500 cm−¹ (red 654 nm) and 350 cm⁻¹ (green 542 nm) as a function of time.

 also monitored with the assistance of a photodetector (shown as PD in Fig. [1\)](#page-3-2). This signal, for the 532 nm pump, is plotted in Fig. [9\(b\)](#page-8-6) (black) showing that, at ∼360 s, a sharp increase in power is observed due to the core being completely filled. 443 This is in agreement with the investigations in Sec. [II C,](#page-4-0) where we show that if the core only is filled with a fluid, visible and near infra-red light is allowed to propagate by index guiding effect. Although it is not clear from Fig. $4(g)$ if 532 nm light is allowed to propagate, experimentally we observed that it 448 can, even though the power levels may be smaller (Fig. $9(b)$, black). This is also true while the cladding capillaries were 450 being filled $(t > 360 \text{ s})$.

 The 532 nm signal shows a 20 s rise time, correspond- ing to the change from almost no guidance to full guidance in the fiber core at 360 s. The 633 nm signal shows a much 454 longer rise time of \sim 70 s (see data in Fig. [5\)](#page-5-1), and hence the filling time of the core can be determined more accurately us- ing a 532 nm pump as this wavelength does not fall within the initial PBG of the fiber. It should be noted that the data in Fig. [5](#page-5-1) is for a 0.5 M solution of glucose with an expected 459 viscosity of 1.26 mPa s, and the data in Fig. $9(b)$ is for a 1 M glucose solution with an expected viscosity of 1.58 mPa s, 461 which results in a longer filling time than shown in Fig. [5.](#page-5-1) This implies that using a different laser source may enhance the ac- curacy of the time measurements of the viscometer. The red trace in Fig. $9(b)$ shows the evolution of the Raman signal at 3500 cm⁻¹ (corresponding wavelength ~654 nm). In this case, no signal is observed before 350 s, just like the PD sig- nal, after which a sharp increase in the count rate was mea-sured, again due to the core being fully filled.

 Once the core is filled and the capillaries continue to fill, the count rate of the inelastically scattered light starts to de- crease as the guiding mechanism changes slowly from index- guiding to a shifted PBG (red trace in Fig. $9(b)$). To deter- mine if the time dependence of the signal is similar over the entire spectral range concerned, the intensity at a shift of at $475 \cdot 350 \text{ cm}^{-1}$ (corresponding wavelength $\sim 542 \text{ nm}$) is plotted in green in Fig. $9(b)$. A lower count rate is observed, but the decay in signal is similar to that at 654 nm. One of the rea- sons for this decay may be that the modal distribution changes when shifting from strong index-guiding, for the core-only filled case, to weakly guidance once fully filled, reducing the 480 power density of the signal, and hence reducing the backscat- ⁴⁸¹ tered signal. 482

Nevertheless, from the results, it is clear that an optimum 483 Raman scatter can be obtained from a fiber when only the core 484 is filled, and when the source is carefully selected to maximize 485 μ the signal. μ

IV. CONCLUSIONS ⁴⁸⁷

We successfully utilized a novel technique to determine 488 when the core and capillaries of a HC-PCF are filled with liquid, and thus determine the average velocity of the liquid, and ⁴⁹⁰ hence its viscosity. The HC-PCF is identified as an ideal vessel for the viscosity measurement of nanoliter quantities of 492 sample, especially important for applications in pharmaceutical and biosciences areas. We analysed the technique for glu- ⁴⁹⁴ cose and fructose solutions in PBS, and demonstrated good ⁴⁹⁵ agreement with values previously published in the literature. ⁴⁹⁶ Our viscometer has the additional advantage of being easily ⁴⁹⁷ integrated to a Raman scattering systems, allowing simulta- ⁴⁹⁸ neous analysis of viscosity and Raman spectral information. ⁴⁹⁹

ACKNOWLEDGMENTS 500

The authors would like to acknowledge Jiadi Lu, Monika $_{501}$ Rutowska, and Andrew Ellis for their support and useful dis- ⁵⁰² cussions. This project is funded by an IRCSET grant, SFI ⁵⁰³ grant 6/IN/I969 and 11/RFP.1/PHY/3233.

¹T. A. Birks, P. J. Roberts, P. S. J. Russell, D. M. Atkin, and T. J. Shepherd, 505 [Electron. Lett.](http://dx.doi.org/10.1049/el:19951306) **31**(22), 1941–1943 (1995). ⁵⁰⁶

2P. Russell, [Science](http://dx.doi.org/10.1126/science.1079280) **299**(5605), 358–362 (2003). ⁵⁰⁷

- ³R. F. Cregan, B. J. Mangan, J. C. Knight, T. A. Birks, P. S. J. Russell, P. J. 508 Roberts, and D. C. Allan, [Science](http://dx.doi.org/10.1126/science.285.5433.1537) **285**(5433), 1537–1539 (1999). ⁵⁰⁹
- 4J. B. Jensen, L. H. Pedersen, P. E. Hoiby, L. B. Nielsen, T. P. Hansen, J. ⁵¹⁰ R. Folkenberg, J. Riishede, D. Noordegraaf, K. Nielsen, A. Carlsen, and A. ⁵¹¹ Bjarklev, [Opt. Lett.](http://dx.doi.org/10.1364/OL.29.001974) **29**(17), 1974–1976 (2004). ⁵¹²

- 6S. Smolka, M. Barth, and O. Benson, [Appl. Phys. Lett.](http://dx.doi.org/10.1063/1.2713132) **90**(11), 111101 ⁵¹⁵ (2007). ⁵¹⁶
- 7J. C. Knight and P. S. J. Russell, [Science](http://dx.doi.org/10.1126/science.1070033) **296**(5566), 276–277 (2002). ⁵¹⁷

⁵M. S. Rutowska, F. C. G. Gunning, F. Kivlehan, E. Moore, D. Brennan, P. ⁵¹³ Galvin, and A. D. Ellis, [Meas. Sci. Technol.](http://dx.doi.org/10.1088/0957-0233/21/9/094016) **21**(9), 094016 (2010). 514

- ⁸ ⁵¹⁸ T. Birks, D. Bird, T. Hedley, J. Pottage, and P. Russell, [Opt. Express](http://dx.doi.org/10.1364/OPEX.12.000069) **12**(1), ⁵¹⁹ 69–74 (2004).
- 9^9 G. Antonopoulos, F. Benabid, T. A. Birks, D. M. Bird, J. C. Knight, and P. ⁵²¹ S. J. Russell, [Opt. Express](http://dx.doi.org/10.1364/OE.14.003000) **14**(7), 3000–3006 (2006).
- 522 ¹⁰ J. M. Fini, [Measurement Sci. Technol.](http://dx.doi.org/10.1088/0957-0233/15/6/011) **15**(6), 1120 (2004).
- ¹¹ ⁵²³ E. C. Bingham, [J. Ind. Eng. Chem. \(Seoul, Repub. Korea\)](http://dx.doi.org/10.1021/ie50063a022) **6**(3), 233–237 524 (1914).
- 525 **(B)**H. Cotton, P. A. Rebers, E. Maudru J, and G. Rorabaugh, *Use of Sug-*⁵²⁶ *ars and Other Carbohydrates in The Food Industry* (American Chemical
- **[Q1](#page-1-0)** ⁵²⁷ Society, 1955), Vol. 12, pp. 3–20. ¹³ ⁵²⁸ P. Arabie and H. Moskowitz, [Atten. Percept. Psychophys.](http://dx.doi.org/10.3758/BF03210240) **9**(5), 410–412
	- ⁵²⁹ (1971). 530 ¹⁴ Y. Komachi, H. Sato, Y. Matsuura, M. Miyagi, and H. Tashiro, [Opt. Lett.](http://dx.doi.org/10.1364/OL.30.002942)
	- ⁵³¹ **30**(21), 2942–2944 (2005).
	- 532 ¹⁵ A. Khetani, V. S. Tiwari, A. Harb, and H. Anis, [Opt. Express](http://dx.doi.org/10.1364/OE.19.015244) 19(16), ⁵³³ 15244–15254 (2011).
	- ¹⁶ K. Nielsen, N. Danny, S. Thorkild, B. Anders, and P. H. Theis, [J](http://dx.doi.org/10.1088/1464-4258/7/8/L02). Opt. A, ⁵³⁵ [Pure Appl. Opt.](http://dx.doi.org/10.1088/1464-4258/7/8/L02) **7**(8), L13 (2005).
	- 536 ¹⁷E. W. Washburn, *[Phys. Rev.](http://dx.doi.org/10.1103/PhysRev.17.273)* **17**(3), 273-283 (1921).
	- 537 ¹⁸ A. H. Pelofsky, [J. Chem. Eng. Data](http://dx.doi.org/10.1021/je60030a031) **11**(3), 394-397 (1966).
	- ¹⁹ ⁵³⁸ E. Aumann, L. M. Hildemann, and A. Tabazadeh, [Atmos. Environ.](http://dx.doi.org/10.1016/j.atmosenv.2009.10.033) **44**(3), 329–337 (2010).
- ²⁰J. F. Comesaña, J. J. Otero, E. García, and A. Correa, [J. Chem. Eng. Data](http://dx.doi.org/10.1021/je020153x) 539 **48**(2), 362–366 (2003). ⁵⁴⁰
- 21A. T. Palasz, J. Thundathil, R. E. Verrall, and R. J. Mapletoft, [Anim. Re-](http://dx.doi.org/10.1016/S0378-4320(99)00090-1) ⁵⁴¹ [prod. Sci.](http://dx.doi.org/10.1016/S0378-4320(99)00090-1) **58**(3–4), 229–24015 (2000). ⁵⁴²
- 22A. A. Ruth, H. Lesche, and B. Nickel, [Z. Phys. Chem.](http://dx.doi.org/10.1524/zpch.217.6.707.20448) **217**(6-2003), 707– ⁵⁴³ $722(2003)$.
- \bigcirc E. Horan, G. Khara, M. Rutowska, A. D. Ellis, and F. C. G. Gunning, 545 $\frac{1}{\sqrt{1-\frac{1$
- 24G. Jones and S. K. Talley, [J. Am. Chem. Soc.](http://dx.doi.org/10.1021/ja01329a024) **55**(2), 624–642 (1933). ⁵⁴⁷
- 25J. Chirife and M. P. Buera, [J. Food Eng.](http://dx.doi.org/10.1016/S0260-8774(97)00060-5) **33**(3–4), 221–226 (1958). ⁵⁴⁸
- 26T. Soesanto and M. C. Williams, [J. Phys. Chem.](http://dx.doi.org/10.1021/j150622a026) **85**(22), 3338–3341 ⁵⁴⁹ (1981). ⁵⁵⁰
- ²⁷R. Mathpal, B. K. Joshi, S. Joshi, and N. D. Kandpal, [Monatsch. Chem.](http://dx.doi.org/10.1007/s00706-005-0435-3) 551 **137**(3), 375–379 (2006). 552
- ²⁸M. Migliori, D. Gabriele, R. Di Sanzo, B. de Cindio, and S. Correra, J. [553](http://dx.doi.org/10.1021/je700062x) [Chem. Eng. Data](http://dx.doi.org/10.1021/je700062x) **52**(4), 1347–1353 (2007). ⁵⁵⁴
- 29V. R. N. Telis, J. Telis-Romero, H. B. Mazzotti, and A. L. Gabas, [Int. J.](http://dx.doi.org/10.1080/10942910600673636) ⁵⁵⁵ [Food Prop.](http://dx.doi.org/10.1080/10942910600673636) **10**(1), 185–195 (2007). ⁵⁵⁶
- 30X. Yang, A. Zhang, D. Wheeler, T. Bond, C. Gu, and Y. Li, [Anal. Bioanal.](http://dx.doi.org/10.1007/s00216-011-5575-1) ⁵⁵⁷ [Chem.](http://dx.doi.org/10.1007/s00216-011-5575-1) **402**(2), 687–691 (2012). ⁵⁵⁸
- 31S. Söderholm, Y. H. Roos, N. Meinander, and M. Hotokka, [J. Raman Spec-](http://dx.doi.org/10.1002/(SICI)1097-4555(199911)30:11<1009::AID-JRS436>3.0.CO;2-) ⁵⁵⁹ [trosc.](http://dx.doi.org/10.1002/(SICI)1097-4555(199911)30:11<1009::AID-JRS436>3.0.CO;2-) **30**(11), 1009–1018 (1999). ⁵⁶⁰